

# Ozone/CLO2 Transmitter

## Model: MDCL100

### Applications

- Various water treatments
- (Bottle washing machine, CIP equipment, rinsing machine)

### Interfered factor

**Ozone:**

- Cl2: OZ7H: it will cause the measured value increasing 1.5%
- OZ7N: Negligible
- ClO2: OZ7N: it will cause the measured value increasing 6

### Description

- **Measurement** Ozone/CLO2
- **Range** 0.00~20.00ppm(mg/l) according to sensor type.
- **Resolution** 0.01ppm(mg/l)
- **Accuracy** ±0.5%±1Digit
- **Storage Temp** -20~70°C
- **Digital filtering** 0~60/ Default...30
- **Calibration Method** 2 Points Calibration, Offset Calibration
- **Relay output** High and Low limit with Dead band setting  
240VAC/0.5A.
- **Current output** 0/4~20mA...2 sets programmable, Max. load 500Ω
- **Max. Pressure** 1.0 bar, without vibration
- **Flow Rate** approx. 30 l/h
- **pH** pH 2 – pH 12
- **Reaction time** T90: approx. 50 sec.
- **Working Temperature:** >5 – <50 °C

### Measured Range

Sensore	Measuring Range(ppm)	Resolution (ppm)
OZ10.12	0.005~2.000	0.001
OZ10.120	0.05~20.00	0.01
CD10.12	0.005~2.000	0.001
CD10.120	0.05~20.00	0.01

### Principle

Film coverage, Ampere dual electrode system

### Temp. compensation:

- Film coverage, Ampere dual electrode system automatically
- compensates through integrated temperature sensors





Address: 7191 Yonge street, Toronto, Canada

Tel: +16472221281(5 line)

Web: [www.madecotech.com](http://www.madecotech.com)

Email: [Info@madecotech.com](mailto:Info@madecotech.com)